

ELISABETH PHARMACON Ltd. Rokycanova 4437/5 615 00 Brno-Zidenice, Czech Republic Phone: +420 542 213 851 E-mail: info@elisabeth.cz Web: www.elisabeth.cz VAT: CZ26258412



# EliGene<sup>®</sup> Soil DNA Preservation Solution Instructions for Use

Package:		Storage:	
Ref. No. 412100PS	Quantity 100 ml	All kit reagents and components should be stored at room temperature (15-30°C). When stored under these conditions, the kit will retain full activity until the expiration date indicated on the kit label.	

# Content

Introduction	2
Equipment Required	2
Kit Contents	
Precautions	2
Protocol	3
Troubleshooting Guide	3



ELISABETH PHARMACON Ltd. Rokycanova 4437/5 615 00 Brno-Zidenice, Czech Republic Phone: +420 542 213 851 E-mail: info@elisabeth.cz Web: www.elisabeth.cz VAT: CZ26258412

## Introduction

**EliGene® Soil DNA Preservation Solution** is intended for storage and transport of soil for DNA or RNA isolation. The solution maintains the viability of bacteria in the soil and at the same time serves as a bacteriostatic to prevent bacterial growth. Nucleic acids are not subject to degradation by DNases or RNases. Samples can be used for cDNA synthesis and 16S rRNA analysis.

Microbial communities are maintained as follows:

- 30 days at temperature 20 °C
- 2 weeks at temperature 4 °C
- 1 week at room temperature
- RNA integrity is maintained for up to 30 days at room temperature

#### The kit is for scientific use only, not for diagnostic use.

#### **Equipment Required**

Centrifuge for 15 ml tubes (2 500 x g) Vortex Pipettes for large volumes

#### **Kit Contents**

Components	Amount
EliGene <sup>®</sup> Soil DNA Preservation Solution	100 ml

### Precautions

Please wear gloves when using this product. Avoid all skin contact with kit reagents. In case of contact, wash thoroughly with water. Do not ingest. See Safety Data Sheets for emergency procedures in case of accidental ingestion or contact.





# Protocol

It is highly recommended to read this information before you use the EliGene® Soil DNA Preservation Solution for the first time.

Add 2 - 2.5 ml of EliGene<sup>®</sup> Soil DNA Preservation Solution per gram of soil sample. In the case
of collecting the soil in tubes with homogenizing beads intended for RNA isolation, add 5 ml of
solution per 2 grams of soil.

Note: When working with sediment samples, use 3 ml of solution per gram of soil.

- 2. Vortex the solution and soil until the sample is completely homogenized. The excess volume of the solution should be in the layer above the homogenized soil.
- 3. Store at 20 °C, 4 °C or room temperature. For RNA processing, centrifuge the sample at 2,500 x g for 5 min. Remove the EliGene<sup>®</sup> Soil DNA Preservation Solution.
- 4. If you were collecting the sample into a homogenizing tube, remove the solution after centrifugation and proceed to the next step for RNA isolation.
- 5. If you have kept the soil in another container, weigh the amount needed and transfer it to a 15 ml homogenizing tube or conical tube and centrifuge to remove the storage solution.

# Troubleshooting Guide

Several types of soil have been tested during the development of the preservation solution. For sediment samples, a larger volume of 3: 1 should be used to dilute the storage solution.

Biomass content plays an important role in soil storage. For soils with unknown microbial culture composition and likely to exceed storage at room temperature, we recommend storing them at - 20 °C or 4 °C. These conditions should ensure that the composition of the microbial community is maintained. More than 2.5 ml of EliGene<sup>®</sup> Soil DNA Preservation Solution per gram of soil may be used. For storage or transport lasting more than 30 days, store samples at -20 °C.



ELISABETH PHARMACON Ltd. Rokycanova 4437/5 615 00 Brno-Zidenice, Czech Republic Phone: +420 542 213 851 E-mail: info@elisabeth.cz Web: www.elisabeth.cz VAT: CZ26258412



# Manufacturer:

ELISABETH PHARMACON Ltd. Rokycanova 4437/5, Brno-Židenice 615 00 info@elisabeth.cz | www.elisabeth.cz | tel.: +420 542 213 851



Catalog number



Batch code



Use by (last day of month)



Upper limit of temperature

Manufacturer

Contains sufficient "N" tests