

ELISABETH PHARMACON Ltd. Rokycanova 4437/5 615 00 Brno-Zidenice, Czech Republic

Phone: +420 542 213 851 E-mail: info@elisabeth.cz Web: www.elisabeth.cz VAT: CZ26258412



# **EliZyme™ LAMP OneS MIX**

#### Intended use:

For Research Use Only. Not for use in diagnostic procedures.

### Storage:

Upon arrival store components at -20 °C. Avoid prolonged exposure to light and keep components on ice when in use. When stored under these conditions, the kit will retain full activity until the expiration date indicated on the kit label. We recommend aliquoting the fluorescent dye at the first use to avoid excess freeze/thaws. Do not store the mix once it is combined with the RTase.

# **Product description**

EliZyme™ LAMP OneS MIX is a revolutionary enzyme system designed to enable fast and sensitive reverse transcription of target RNA, followed by isothermal amplification, all in a single tube.

The kit employs the power of Bst Polymerase, which is provided in a convenient 2X mix format. This polymerase possesses exceptional strand displacement activity, enabling DNA synthesis to occur at a consistent temperature without the need for thermal cycling. Derived from the large fragment of Geobacillus stearothermophilus (formerly Bacillus stearothermophilus) DNA Polymerase, this specific part of the protein is responsible for catalyzing the synthesis of DNA in the 5'-3' direction, but lacks the 5'-3' exonuclease domain.

Reverse transcription of target RNA is accomplished through the utilization of the thermostable and highly active RTase, which is combined with an RNase inhibitor to protect the RNA from degradation caused by contaminating RNase enzymes.

EliZyme™ LAMP OneS MIX incorporates an advanced buffer system that ensures excellent yield and performance, even in challenging conditions such as the presence of inhibitors. The kit only requires the addition of primers, template, and water, and it includes a separate fluorescent dye for real-time detection using any qPCR instrument.

#### Content

|                    | Ref. No. | Content   | Size     |
|--------------------|----------|---|----------|
| EliZyme™ LAMP OneS |          | $1\times1.25$ ml $2\times$ mix + $1\times125$ $\mu$ l |          |
| MIX                | EZ2901   | 20× Dye + 1×200 μl RTase                              | 100 rxns |
|                    |          | 4×1.56 ml 2× mix + 1×625 μl                           |          |
|                    | EZ2905   | 20× Dye + 1×1 ml RTase                                | 500 rxns |



ELISABETH PHARMACON Ltd.
Rokycanova 4437/5
615 00 Brno-Zidenice, Czech Republic

Phone: +420 542 213 851 E-mail: info@elisabeth.cz Web: www.elisabeth.cz VAT: CZ26258412



| EliZyme™ LAMP OneS MIX | 2× mix | 6 mM MgSO <sub>4</sub> , 3.2 mM dNTPs |  |
|------------------------|--------|---------------------------------------|--|
|------------------------|--------|---------------------------------------|--|

The buffers composition has been optimized to maximize amplification success rates.

#### **LAMP**

Incubate at 65 °C for 30 minutes. Time can be extended and temperature can be modified (between 55 °C and 70 °C) as necessary for low copy targets, challenging templates, or whenever amplification times have been reported to be slow.

To improve sensitivity and specificity we recommend a quick cycle at the beginning of the reaction to facilitate the correct annealing of primers to the RNA template (for example 65 °C for 1 min, followed by a slow reduction of temperature to ambient value before starting the amplification).

If a qPCR instrument is used for signal detection, follow the reaction using the FAM channel, acquiring data every 10-15 seconds. If final products are to be analyzed after the reaction is complete, the enzyme can be inactivated by heating at 80 °C for 10 minutes.

#### **Primers**

We recommend a predicted melting temperature of around 60  $^{\circ}$ C using default Primer Explorer v5 settings. A primer set can be prepared with all 4 or 6 (if you include Loop) primers. A 10X primer set should contain: 16µM FIP, 16µM BIP, 2µM F3, 2µM B3, 4-8µM LoopF, 4-8µM LoopB in TE Buffer or water.

#### Reaction setup

## EliZyme™ LAMP OneS MIX

After thawing, briefly vortex all components and shortly spin. Reactions should be setup on ice.

| Reagent                      | 25 μl reaction | Final conc. |
|------------------------------|----------------|-------------|
| 2× EliZyme™ Bst Mix          | 12.5 μΙ        | 1×          |
| 20× EliZyme™ Fluorescent Dye | 1.25 μl        | 1×          |
| RTase                        | 2 μΙ           | 1×          |
| 10× Primer set               | 2.5 μl         | 1×          |
| Template DNA                 | Variable       |             |
| PCR grade water              | Up to 25 μl    |             |

## LAMP profile

| Step                | Temperature | Time      |
|---------------------|-------------|-----------|
| Strand displacement | 55-70 °C*   | 30-60 min |
| Deactivation**      | 80 °C       | 10 min    |

<sup>\*</sup>Standard incubation at 65 °C for 30 minutes.

<sup>\*\*</sup>If final products are to be analyzed after the reaction is complete, the enzyme can be



ELISABETH PHARMACON Ltd.

Rokycanova 4437/5 615 00 Brno-Zidenice, Czech Republic

Phone: +420 542 213 851 E-mail: info@elisabeth.cz Web: www.elisabeth.cz VAT: CZ26258412



# Manufacturer:

# ELISABETH PHARMACON, spol. s r. o.

Rokycanova 4437/5, Brno-Židenice 615 00

info@elisabeth.cz | www.elisabeth.cz | tel.: +420 542 213 851



Catalog number



Batch code



Use by (last day of month)



Upper limit of temperature



Manufacturer



Contains sufficient "N" tests

Version: 240823-01 Page **3** of **3**